



Instructions for Use

non-pTAU ELISA



Rev. 0/2023

For research use only

Order Number:

847-0108000102-RUO	96 reactions
--------------------	--------------

This documentation describes the state at the time of publishing. It needs not necessarily agree with future versions.
Subject to change!

Print-out and further use permitted with indication of source.

© Copyright 2023, Roboscreen GmbH

Manufacturer:

Roboscreen GmbH	Phone: +49 341 989734 0	www.roboscreen.com
Hohmannstraße 7	Fax: +49 341 989734 199	info@roboscreen.com
04129 Leipzig		
Germany		

Publication No.: Manual_non-pTAU ELISA_RUO_e_rev0	0/2023
--	--------

Contents

1	Introduction	4
1.1	Intended use	4
1.2	Warranty and technical support	4
1.3	Notes on the use of this instructions for use	5
2	Safety precautions.....	6
3	Test principle	8
4	Performance assessment	9
5	Kit components	10
6	Preparation of components.....	12
6.1	1X Wash solution	12
6.2	Standards D3.1-D3.6.....	12
6.3	Controls D7 and D8.....	12
6.4	1X HRP conjugate	13
7	Storage and expiry date.....	14
8	Components not included in the kit	14
9	Procedure notes	15
10	Specimen collection and storage	16
10.1	Specimen collection.....	16
10.2	Specimen storage	16
10.3	Specimen dilution	17
11	Test procedure	18
12	Data analysis	20
12.1	Quality criteria of the assay	20
12.2	Calculation of unknown concentration of tau w/o phosphorylation at T175 or T181.....	20

1 Introduction

1.1 Intended use

The non-pTAU ELISA is an enzyme immunoassay intended for the quantitative determination of a tau fraction without (w/o) phosphorylation at T175/T181 in human cerebrospinal fluid (CSF). All contents of the non-pTAU ELISA are produced under the guidelines of quality control accordingly to the DIN EN ISO 13485 requirements.

The non-pTAU ELISA is for research use only and is not intended as a diagnostic test.

1.2 Warranty and technical support

The manufacturer guarantees the correct functioning of the kit for the applications described in the instructions for use (IFU). During the warranty period, non-pTAU ELISA allows for precise and reproducible data collection in connection with superior sensitivity. Any warranty claims shall only be valid if the general principles of Good Laboratory Practice (GLP) and the manufacturer's recommendations are observed.

To improve the application and design, Roboscreen GmbH reserves the right of product replacement or modification. The manufacturer may be contacted at any time for questions and problems or technical support concerning the quantification of tau w/o phosphorylation T175/T181 in CSF.


CONSULT INSTRUCTIONS FOR USE



This package insert must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of results cannot be guaranteed if there are any deviations from the instructions in this package insert.

1.3 Notes on the use of this instructions for use

For easy reference and orientation, the IFU uses the following warning and information symbols as well as the shown methodology:

	REF Catalogue number
	Content Contains sufficient reagents for <N> tests
	Storage conditions
	Consult instructions for use
	Expiration date
	Manufactured by
	For single use only

The following abbreviations are used in the IFU:

CSF	Cerebrospinal fluid
CV	Coefficient of variation
ELISA	Enzyme-linked immunosorbent assay
GLP	Good Laboratory Practice
HRP	Horseradish peroxidase
OD	Optical density
RT	Room temperature (18-25°C)
TMB	Tetramethylbenzidine

2 Safety precautions

We recommend reading this chapter thoroughly before using this kit, to ensure the safety of the user and error-free utilization.

Any safety instructions and additional information of this IFU must be observed at all times.

Read and make sure you understand the operating instructions completely and thoroughly before carrying out the test. Use the currently valid version from the kit.

Notify the respective supplier in writing within one week from receiving the merchandise, should the test pack be substantially damaged. Damaged components must not be used to carry out the assay, however, they should be kept until the transport damages are finally settled.

Comply with Good Laboratory Practice and safety regulations. Wear laboratory coats, disposable Latex gloves and safety goggles whenever the need arises.

Reagents of this kit which contain hazardous substances may cause irritations to eyes and skin. See indications under COMPONENTS OF THE KIT and on the labels. Safety data sheets of this product are available upon request.

Chemicals and prepared or used reagents shall be disposed of as hazardous waste in compliance with the respective national regulations.

The cleaning staff has to be instructed by experts with regard to any potential risks and the appropriate handling of such substances.

Avoid any contact with stopping solution. This may cause irritations to the skin and chemical burns.



FOR SINGLE USE ONLY!

This kit is made for single use only!


ATTENTION!

Don't eat or drink components of the kit!

The kit shall only be handled by educated personnel in a laboratory environment!

3 Test principle

This kit works by means of monoclonal antibody 1G2 that specifically recognizes tau protein w/o phosphorylation at position T175 and T181, immobilized on the surface area of the microtiter plate. Tau protein w/o phosphorylation at T175/T181 from samples, standards and controls is trapped by this antibody in presence of another peroxidase conjugated monoclonal anti-tau antibody that specifically binds to amino acids 155-165 of human tau protein. Amount of bound conjugated antibody is estimated using chromogenic substrate tetramethylbenzidine (TMB). The concentration of tau protein w/o phosphorylation at T175/T181 is proportional to the obtained optical density.



1. Ready-to-use: Capture antibody coated on well plate
2. Binding of target antigen simultaneously by capture antibody and HRP-conjugated antibody
3. Direct detection via HRP-conjugated antibody and peroxide/TMB.

4 Performance assessment

The table below shows typical data for calibration curves. Do not use for calculation!

Standard	Tau (pg/ml)	3 h Incubation 150 rpm* (21,5 ± 3,5 °C)	
		OD _{mean}	OD/OD _{max} (%)
D3.1	1200	2.682	100
D3.2	600	1.815	68
D3.3	120	0.572	21
D3.4	50	0.409	15
D3.5	25	0.294	11
D3.6	12	0.123	5

* Shakers have different forces due to their specifications (deflection in mm) at a frequency of 150 rpm (see note page 18).

5 Kit components

Component		96	Description
Immunostrips D1		12 x 8	Coated immunostrips containing 1G2 antibody, blocked and stabilized. Ready to use.
40X Wash buffer D2		1 x 50 ml	40X Wash buffer containing PBS, detergent and proclin 300.
Standards		6 x 3	Dried tau standards (STD) for preparing a standard curve for quantification of TAU w/o phosphorylation in unknown cerebrospinal fluid samples. Containing PBS, protein and proclin 300.
D3.1		3	1200 pg tau
D3.2		3	600 pg tau
D3.3		3	120 pg tau
D3.4		3	50 pg tau
D3.5		3	25 pg tau
D3.6		3	12 pg tau
30X HRP conjugate D5		1 ml	Monoclonal anti-Tau antibody conjugated with horseradish peroxidase, 30X concentrate containing PBS, protein, detergent and proclin 300.
Assay buffer D6		50 ml	Assay buffer containing PBS, protein, detergent and proclin 300. Ready to use.

Component		96	Description
Control high D7		3	Dried Tau high positive control (CTRL), containing PBS, protein and proclin 300.
Control low D8		3	Dried Tau low positive control (CTRL), containing PBS, protein and proclin 300.
Staining solution D9		20 ml	TMB/peroxide solution. Ready to use.
Stop solution D10		25 ml	1 M sulphuric acid. Ready to use.
Sealing tape		1	
Instructions for use		1	

6 Preparation of components

6.1 1X Wash solution

Dilute 40X wash buffer D2 using de-ionized or bi-distilled water before the first wash step of the immunoassay.

Volume of 1X Wash solution	Volume of 40X Wash buffer D2	Volume of de-ionized or bi-distillated water
400 ml	10 ml	390 ml
600 ml	15 ml	585 ml
800 ml	20 ml	780 ml
1000 ml	25 ml	975 ml

6.2 Standards D3.1-D3.6

Add 0.5 ml of assay buffer D6 to each standard vial D3.1 – D3.6 and mix quickly, e.g. within 2 s by vortex.

6.3 Controls D7 and D8

Add 0.5 ml of assay buffer D6 to each control vial D7 and D8 and mix quickly, e.g. within 2 s by vortex.

6.4 1X HRP conjugate

Dilute 30X HRP conjugate D5 at ratio 1:15* with assay buffer D6. Mix by means of shaking the tube.

Number of immunostrips	Volume of 30X HRP D5	Volume of assay buffer D6
1 – 4	0.25 ml	3.50 ml
5 – 8	0.40 ml	5.60 ml
9 - 12	0.60 ml	8.40 ml

* Dilution of the conjugate is 1:15. In the test, the conjugate is present at a dilution of 1:30 by the addition of 50 µl standard / control / sample to 50 µl 1:15 diluted HRP conjugate.

7 Storage and expiry date

The kit is delivered at ambient temperature and should be stored at $6 \pm 4^{\circ}\text{C}$. Protect from heat and direct sunlight. Under these conditions, the kit has a life time as indicated on the kit box while retaining its endurance and stability.

Prepared kit components have the following expiry dates:

Component	Preparation step	Expiry date
D1	Coated immunostrips after opening of the bag, taking out strips and closing the bag.	At $6 \pm 4^{\circ}\text{C}$ up to 4 weeks.
D2	1X Ready-to-use washing solution.	At $6 \pm 4^{\circ}\text{C}$ up to 1 week.
D3.1-D3.6	Standards D3.1-D3.6 dissolved in D6.	At $6 \pm 4^{\circ}\text{C}$ up to 4 h.
D7, D8	Controls D7 and D8 dissolved in D6.	At $6 \pm 4^{\circ}\text{C}$ up to 4 h.
D5	Ready-to-use HRP-conjugate 1:15 diluted.	At $6 \pm 4^{\circ}\text{C}$ up to 4 h.

8 Components not included in the kit

- Calibrated micropipettes with CV < 3%,
Volume: 10-100 μl ; 100-1000 μl .
- 8-channel micropipette with reagent reservoirs.
- Plate shaker* 100-1500 rpm e.g. Rotamax 120.
- Vortex mixer.
- Automated or semi-automated ELISA plate washing system.
- Bi-distilled or de-ionized water.
- Paper towels, pipette tips and timer.
- ELISA plate reader for reading absorbance at 450 and 620 nm.
- Polypropylene tubes for sample dilution.

*Shakers have different forces due to their specifications (deflection in mm) at a frequency of 150 rpm (see note on page 17).

9 Procedure notes

Any improper handling of samples or modification of the test procedure may influence the results. The indicated volumes, incubation times, temperatures and pretreatment steps must be followed strictly regarding this instruction.

Be sure that required reagents, materials and devices are prepared ready at the appropriate time. Allow all staining solution D9 to reach room temperature (21.5 ± 3.5 °C). Mix assay buffer D6 and 30X HRP conjugate D5 by vortex before use.

Avoid contamination of reagents, pipettes and immunostrips/tubes by use of different disposables between different samples and components. Do not interchange caps. Do not re-use any well, tube or reagent.

All measurements can be done in **single determinations**; however a dual determination increases the safety of the results and allows additional evaluations for the precision of the measurements.

For duplicate determinations, R^2 of the standard curve should be ≥ 0.99 . It is recommended to use a pipetting scheme to apply all STD, CTRL and samples. Solution of 1X HRP conjugate **D5**, staining solution D9 and stop solution D10 should be transferred by a 8-channel micropipette or a pipette with reservoir (multistep pipette) to all wells of the immunostrips.

Washing should be done by 8-channel micropipette or ELISA plate washer. Avoid drying and over stressing of wells and control exact washing of all wells.

10 Specimen collection and storage

10.1 Specimen collection

Lumbar puncture may be performed at the vertebral body L3-L5 with the patient either sitting or lying down. Use a small diameter (0.7 mm and 22 G), preferably not traumatic needle. A small-gauge needle will make a smaller hole in the Dura mater, aiding healing. Usage of a non-traumatic needle will reduce the chance of blood contamination in the CSF.

Each laboratory should use one kind of polypropylene tubes only. Glass or polystyrene tubes should in no circumstances be used. Tubes of the smallest volume should be used, and these should be filled to at least 50% of their volume. It is important to have carefully recorded and validated details concerning each stored sample so that any investigator when using these samples has a precise history of the sample.

Centrifugation is only required for visually hemorrhagic samples. Centrifuge immediately with recommended 2000 x g at RT for 10 min.

10.2 Specimen storage

It is recommended to freeze samples and store at -80°C for long time storage. It is recommended to limit the number of freeze /thaw cycles to a maximum of 1-2. Samples should be stored no longer than 2 years.

Note

For dilution of CSF use polypropylene tubes or dilute directly onto immunostrips D1.

10.3 Specimen dilution

Samples showing an OD higher than OD of highest standard D3.1 should be diluted before test procedure using assay buffer D6.

11 Test procedure

1. Transfer 50 µl of 1X HRP conjugate in each well.
2. Afterwards pipette 50 µl of each standard, control and patient sample into the respective wells of plate.
3. The sequence of pipetting steps can be reversed.
4. Mix thoroughly e.g. by pipette at least 5 times.
5. Cover plate with lid or foil.
6. Incubate immunoplate for 3 h ± 15 min at 21.5 ± 3.5 °C with 150 ± 15 rpm at RT.

NOTE

Shakers have different forces due to their deflection at a frequency of 150 rpm. The calculation of the acceleration (a) according to the formula $a = 4 \pi^2 r n^2$ should give a value of 2.5 m / s².

Radius (mm)	Number of revolutions (rpm)
10	150
5	212
3	274
1,5	387
0.5	671

$$\pi^2 = 9.87, r \text{ in m (10 mm = 0.01 m) and } n \text{ in r/s (150 rpm/60 = 2.5 r/s)}$$

7. Wash plate 5 x with 300 µl/well of 1X wash solution using an automatic ELISA plate washer.

Note

Alternatively, when performed manually, discard incubation solution. Remove excess solution after washing by tapping immunostrip on paper towel.

8. Pipette 100 µl of staining solution D9 into each well.
9. Incubate plate at RT in the dark for 30 min.
10. Stop the substrate reaction by adding 150 µl of stop solution D10 into each well. Briefly mix contents in the plate reader.
11. Measure optical density with a photometer at 450 nm using 620 nm as reference wave length within 15 min after stopping.

Note

In samples with a high concentration of tau protein formed dye may be precipitated due to intensive staining. Therefore, 15 min time lapse until the measurement takes place are recommended.

12 Data analysis

12.1 Quality criteria of the assay

- Concentration of positive control **D7** and **D8** should be inside range corresponding to lot specific certificate of analysis.

12.2 Calculation of unknown concentration of tau w/o phosphorylation at T175 or T181

For the determination of the non-phosphorylated tau concentration in controls and samples the automatic data analysis by means of reader software, usually the logistic regression with 4 or 5 parameters or logit-log method is recommended. The standard curve typically shows a linear progression between the plateau of the highest standard D3.1 (1200 pg/ml) and the lowest standard D3.6 (12 pg/ml).

Note

Samples with a measured OD smaller than the OD of the lowest standard D3.6 can be reported in terms of non-phosphorylated Tau protein concentration <12 pg/ml.
